



- NOTES**
- ALL WIRES 20-99 #18 AWG.
 - WIRE COLOR CODE:
RED SOLID or STRIPE - 24 Vac CONTROL
BLUE SOLID or BLUE WITH WHITE or GRAY STRIPE - 24 Vac NEUTRAL (COMMON)
PURPLE SOLID - 24 Vdc
WHITE SOLID - 0 Vdc (COMMON)
BROWN SOLID - 117 Vac POWER
BLUE WITH BLACK STRIPE - 117 Vac NEUTRAL (COMMON)
ALL 2- CONDUCTOR CABLES 24 V
 - REQUIRED FOR SENSLAS OPTION.
 - REQUIRED FOR DUAL GAS OPTION.

REV	DESCRIPTION	DATE	APPROVED	REV	DESCRIPTION	DATE	APPROVED
4	MOVED 33-4 GAS COMMON CONNECTION TO DWIG 802-101780	10/30/13	S BARBER	4	MOVED 33-4 GAS COMMON CONNECTION TO DWIG 802-101780	10/30/13	S BARBER
3	ADDED WIRE 67 RED AND CHANGED TERMINATION OF WIRE 60	05/31/13	S BARBER	3	ADDED WIRE 67 RED AND CHANGED TERMINATION OF WIRE 60	05/31/13	S BARBER
2	ADDED SENSLAS OPTION CONNECTIONS	05/20/13	S BARBER	2	ADDED SENSLAS OPTION CONNECTIONS	05/20/13	S BARBER
1	REVISED GAS 1, GAS 2 & SW9-X1 CONNECTIONS; DELETED DS7	05/11/13	S BARBER	1	REVISED GAS 1, GAS 2 & SW9-X1 CONNECTIONS; DELETED DS7	05/11/13	S BARBER

APPROVALS	DATE	FILE
DRAWN S BARBER	04/25/2013	CONTROL CONSOLE LA-306
CHECKED S BARBER	04/25/2013	Document Number
ISSUED J CLARK	05/12/2013	Size STD 802-101784-02

LCI Furnaces
 DIVISION OF LOCHABER CORNWALL, INC.
 675 NORTH ECKHOFF STREET, BLDG D, ORANGE, CA 92668 USA

Control Console LA-306
 Date: Saturday, March 01, 2014
 Sheet 1 of 3